EU DECLARATION OF CONFORMITY

Fresenius Medical Care AG & Co. KGaA

Else-Kröner-Str. 1
61352 Bad Homburg – Germany

SRN: DE-MF-000008193

declares under his sole responsibility that the product

Product Name:	*multiFiltratePRO paed HD*
GMN / Basic UDI-DI:	*4039361-0000-0000-0117-NQ*
Product group:	*Acute dialysis tubing sets multifiltrate*
EMDN Code	*F020102*
Intended Purpose:	*Channeling of blood and fluid in an extracorporeal treatment*

meets the applicable provisions as stated below and the relevant standards and common specifications as specified in the technical documentation.

Applicable regulation/s:	*European Medical Device Regulation 2017/745*
Risk class: (according to Annex VIII Medical Device Regulation 2017/745)	*Class IIa according to Rule 2*
Conformity assessment procedure:	*Annex IX (Chapters I and III) of the Regulation (EU) 2017/745 (MDR)*
Notified body:	*TÜV SÜD Product Service GmbH* *Ridlerstraße 65* *80339 München – Germany*
Notified body no.:	*0123*
EU certificate:	*G10 077174 0005*
Place, Date:	*Bad Homburg, 14-Apr-2022*

.

i.V. _____
Dr. Rüdiger Amberg
Product Center Responsible Person

.

ppa _____
Dr. Thomas Himstedt
Responsible person acc. MDR Art. 15

This Declaration of Conformity is valid until withdrawn or reissued due to significant product change, new product code or new/changed regulatory/legal requirement.

EU DECLARATION OF CONFORMITY

Versions:

Product code	Product name
F00008936	multiFiltratePRO paed HD

Accessories:

(according to the European Medical Device Regulation 2017/745)

Product code	Product name
-	-

This Declaration of Conformity is valid until withdrawn or reissued due to significant product change, new product code or new/changed regulatory/legal requirement.