

# EU DECLARATION OF CONFORMITY

## Fresenius Medical Care AG

Else-Kröner-Str. 1  
61352 Bad Homburg – Germany  
SRN: DE-MF-000008193

**declares under his sole responsibility that the product**

Product Name: *multiFiltratePRO*  
Product Code: *M205001*  
GMN / Basic UDI-DI: *4039361-0000-0000-0006-N9*  
Product Group: *multi*  
EMDN Code: *Z120902*  
Intended Purpose: *Control, operation and monitoring of extracorporeal treatment*  
Software Version *V6.0*

**meets the applicable provisions as stated below and the relevant standards and common specifications as specified in the technical documentation.**

Applicable regulation/s: *- European Medical Device Regulation 2017/745  
- European Directive 2011/65/EU Restriction of Hazardous Substances*

Risk class: *Class IIb according to Rule 12*  
(according to Annex VIII Medical Device Regulation 2017/745)

Conformity assessment procedure: *Annex IX (Chapters I and III) of the Regulation (EU) 2017/745 (MDR)*


Notified body: *TÜV SÜD Product Service GmbH*  
(according to Medical Device Regulation 2017/745) *Ridlerstraße 65  
80339 München – Germany*

Notified body no.: *0123*

EU certificate: *G10 077174 0005*

Place, Date: *Bad Homburg, 01-Dec-2023*

i.V.   
Daniel Pelzi  
Product Center Responsible Person

ppa   
Dr. Thomas Himstedt  
Responsible person acc. MDR Art. 15