

EU DECLARATION OF CONFORMITY

Fresenius Medical Care AG

Else-Kröner-Str. 1
61352 Bad Homburg – Germany
SRN: DE-MF-000008193


declares under his sole responsibility that the product

Product Name: *PatientOnLine*
Product Code: *F989901*
GMN / Basic UDI-DI: *4039361-0000-0000-0010-MU*
Product Group: *PatientOnLine (POL)*
EMDN Code: *Z12099092*
Intended Purpose: *Acquire, calculate, model, save, display and transfer parameters of dialysis treatments*
Software-Version: *6.4*

meets the applicable provisions as stated below and the relevant standards and common specifications as specified in the technical documentation.

Applicable regulation/s: *European Medical Device Regulation 2017/745*
Risk class: *Class IIa according to Rule 11*
(according to Annex VIII Medical Device Regulation 2017/745)
Conformity assessment procedure: *Annex IX (Chapters I and III) of the Regulation (EU) 2017/745 (MDR)*
Notified body: *TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München – Germany*
Notified body no.: *0123*
EU certificate: *G10 077174 0005*
Place, Date: *Bad Homburg, 01-Dec-2023*

i.V. 
Daniel Pelzl
Product Center Responsible Person

ppa 
Dr. Thomas Himstedt
Responsible person acc. MDR Art. 15

EU DECLARATION OF CONFORMITY

Fresenius Medical Care AG

Else-Kröner-Str. 1
61352 Bad Homburg – Germany
SRN: DE-MF-000008193


declares under his sole responsibility that the product

Product Name: *PatientOnLine Multi-User*
Product Code: *F989951*
GMN / Basic UDI-DI: *4039361-0000-0000-0010-MU*
Product Group: *PatientOnLine (POL)*
EMDN Code: *Z12099092*
Intended Purpose: *Acquire, calculate, model, save, display and transfer parameters of dialysis treatments*
Software-Version: *6.4*

meets the applicable provisions as stated below and the relevant standards and common specifications as specified in the technical documentation.

Applicable regulation/s: *European Medical Device Regulation 2017/745*
Risk class: *Class IIa according to Rule 11*
(according to Annex VIII Medical Device Regulation 2017/745)
Conformity assessment procedure: *Annex IX (Chapters I and III) of the Regulation (EU) 2017/745 (MDR)*
Notified body: *TÜV SÜD Product Service GmbH
Ridlerstraße 65
80339 München – Germany*
Notified body no.: *0123*
EU certificate: *G10 077174 0005*
Place, Date: *Bad Homburg, 01-Dec-2023*

i.V. 
Daniel Pelzl
Product Center Responsible Person

ppa 
Dr. Thomas Himstedt
Responsible person acc. MDR Art. 15